


**TEST REPORT**  
**07-2017-BAL-212Met**

**Sample No.** : Met-447  
**Type of Job** : ON-SITE CALIBRATION  
**Date Calibrated** : July 17, 2017  
**Sample** : Table Balance  
**Manufacturer** : China  
**Serial No.** : 0014  
**Model** : JPT-10  
**Graduation** : 1 g  
**Capacity** : 1000 g  
**Classification of Balance** : Class III (Ordinary Accuracy Class)  
**Company** : **SURIGAO STATE COLLEGE OF TECHNOLOGY**  
**Address** : Narciso Street, Surigao City  
**Page** : 1 of 3


This instrument was calibrated using reference standard traceable to SI Units as maintained by the National Metrology Institute- ITDI, Philippines. The following results were obtained:

**I. Shift Test at 500 g**

Position	Difference (g)	Maximum Permissible Error (MPE)
(1-2)	0.0	$\pm 3.0$ g
(1-3)	0.0	$\pm 3.0$ g
(1-4)	0.0	$\pm 3.0$ g
(1-5)	0.0	$\pm 3.0$ g



OP-026-F14  
Revision 1



**II. Repeatability Test at Half and Full Load**

Load (g)	Difference (g)	Maximum Permissible Error (MPE)
500	0.0	± 3.0 g
1000	0.0	± 3.0 g


**III. Departure from Nominal Value (Increasing & Decreasing Load)**

Load (g)	Increasing Load Reading (g)	Error (g)	Decreasing Load Reading (g)	Error (g)	Maximum Permissible Error (MPE)	Uncertainty of Measurement (±g)
10	10.0	0.0	10.0	0.0	+ 1.0 g	0.38
20	20.0	0.0	20.0	0.0	+ 1.0 g	0.38
50	50.0	0.0	50.0	0.0	+ 1.0 g	0.38
100	100.0	0.0	100.0	0.0	+ 2.0 g	0.38
200	200.0	0.0	200.0	0.0	+ 2.0 g	0.38
500	500.0	0.0	500.0	0.0	+ 3.0 g	0.38
750	751.0	1.0	751.0	1.0	+ 3.0 g	0.38
1000	1001.0	1.0	1001.0	1.0	+ 3.0 g	0.38

Environmental Conditions : Relative Humidity : 54.0%  
 Ambient Temperature : 25.5°C

Counter Weights	Number of Holes	Number of Leaded Holes
10 g	-	-
20 g (1)	-	-
20 g (2)	-	-
50 g	-	-
100 g	-	-
200 g (1)	-	-
200 g (2)	-	-
500 g	-	-






**TEST REPORT**  
**07-2017-BAL-212Met**

**Sample No.** : Met-448  
**Type of Job** : ON-SITE CALIBRATION  
**Date Calibrated** : July 17, 2017  
**Sample** : Table Balance  
**Manufacturer** : China  
**Serial No.** : 0013  
**Model** : JPT-10  
**Graduation** : 1 g  
**Capacity** : 1000 g  
**Classification of Balance** : Class III (Ordinary Accuracy Class)  
**Company** : **SURIGAO STATE COLLEGE OF TECHNOLOGY**  
**Address** : Narciso Street, Surigao City  
**Page** : 1 of 3


This instrument was calibrated using reference standard traceable to SI Units as maintained by the National Metrology Institute- ITDI, Philippines. The following results were obtained:

**I. Shift Test at 500 g**

Position	Difference (g)	Maximum Permissible Error (MPE)
(1-2)	0.0	$\pm 3.0$ g
(1-3)	0.0	$\pm 3.0$ g
(1-4)	0.0	$\pm 3.0$ g
(1-5)	0.0	$\pm 3.0$ g



OP-026-F14  
Revision 1



## II. Repeatability Test at Half and Full Load

Load (g)	Difference (g)	Maximum Permissible Error (MPE)
500	0.0	$\pm 3.0$ g
1000	0.0	$\pm 3.0$ g


## III. Departure from Nominal Value (Increasing & Decreasing Load)

Load (g)	Increasing Load Reading (g)	Error (g)	Decreasing Load Reading (g)	Error (g)	Maximum Permissible Error (MPE)	Uncertainty of Measurement ( $\pm$ g)
10	10.0	0.0	10.0	0.0	$\pm 1.0$ g	0.38
20	20.0	0.0	20.0	0.0	$\pm 1.0$ g	0.38
50	50.0	0.0	50.0	0.0	$\pm 1.0$ g	0.38
100	100.0	0.0	100.0	0.0	$\pm 2.0$ g	0.38
200	200.0	0.0	200.0	0.0	$\pm 2.0$ g	0.38
500	500.0	0.0	500.0	0.0	$\pm 3.0$ g	0.38
750	750.0	0.0	750.0	0.0	$\pm 3.0$ g	0.38
1000	1000.0	0.0	1000.0	0.0	$\pm 3.0$ g	0.38

Environmental Conditions : Relative Humidity : 50.0%  
Ambient Temperature : 25.3°C

Counter Weights	Number of Holes	Number of Leaded Holes
10 g	-	-
20 g (1)	-	-
20 g (2)	-	-
50 g	-	-
100 g	-	-
200 g (1)	-	-
200 g (2)	-	-
500 g	-	-







**IV. Remarks:**

1. The uncertainty of measurement is estimated at 95% level of confidence with a coverage factor  $k=2$ .
2. The above values are those obtained at the time of test and refer only to the particular instrument calibrated.
3. The end-user shall determine the suitability of this instrument for its intended use.
4. This report shall not be reproduced in any form, except in full, without written approval of the laboratory.

Calibrated by:


  
**ENGR. MANOLITO R. TAPANGAN**  
Laboratory Analyst

Reviewed by:

  
**GIDEON M. TANGHAL**  
Laboratory Analyst

Certified Correct and  
Approved for Release by:

  
**JENNIFER J. DEJARME**  
Chief Laboratory Analyst




**TEST REPORT**  
**07-2017-BAL-212Met**

**Sample No.** : Met-449  
**Type of Job** : ON-SITE CALIBRATION  
**Date Calibrated** : July 17, 2017  
**Sample** : Table Balance  
**Manufacturer** : China  
**Serial No.** : 0012  
**Model** : JPT-10  
**Graduation** : 1 g  
**Capacity** : 1000 g  
**Classification of Balance** : Class III (Ordinary Accuracy Class)  
**Company** : **SURIGAO STATE COLLEGE OF TECHNOLOGY**  
**Address** : Narciso Street, Surigao City  
**Page** : 1 of 3

This instrument was calibrated using reference standard traceable to SI Units as maintained by the National Metrology Institute- ITDI, Philippines. The following results were obtained:


**I. Shift Test at 500 g**

Position	Difference (g)	Maximum Permissible Error (MPE)
(1-2)	0.0	$\pm 3.0$ g
(1-3)	0.0	$\pm 3.0$ g
(1-4)	0.0	$\pm 3.0$ g
(1-5)	0.0	$\pm 3.0$ g



OP-026-F14  
Revision 1





## II. Repeatability Test at Half and Full Load


Load (g)	Difference (g)	Maximum Permissible Error (MPE)
500	0.0	$\pm 3.0$ g
1000	0.0	$\pm 3.0$ g

## III. Departure from Nominal Value (Increasing & Decreasing Load)

Load (g)	Increasing Load Reading (g)	Error (g)	Decreasing Load Reading (g)	Error (g)	Maximum Permissible Error (MPE)	Uncertainty of Measurement ( $\pm$ g)
10	10.0	0.0	10.0	0.0	$\pm 1.0$ g	0.38
20	20.0	0.0	20.0	0.0	$\pm 1.0$ g	0.38
50	50.0	0.0	50.0	0.0	$\pm 1.0$ g	0.38
100	100.0	0.0	100.0	0.0	$\pm 2.0$ g	0.38
200	200.0	0.0	200.0	0.0	$\pm 2.0$ g	0.38
500	500.0	0.0	500.0	0.0	$\pm 3.0$ g	0.38
750	750.0	0.0	750.0	0.0	$\pm 3.0$ g	0.38
1000	1000.0	0.0	1000.0	0.0	$\pm 3.0$ g	0.38

Environmental Conditions : Relative Humidity : 49.0%  
Ambient Temperature : 25.6°C


Counter Weights	Number of Holes	Number of Leaded Holes
10 g	-	-
20 g (1)	-	-
20 g (2)	-	-
50 g	-	-
100 g	-	-
200 g (1)	-	-
200 g (2)	-	-
500 g	-	-



**IV. Remarks:**

1. The uncertainty of measurement is estimated at 95% level of confidence with a coverage factor  $k=2$ .
2. The above values are those obtained at the time of test and refer only to the particular instrument calibrated.
3. The end-user shall determine the suitability of this instrument for its intended use.
4. This report shall not be reproduced in any form, except in full, without written approval of the laboratory.


Calibrated by:

  
**ENGR. MANOLITO R. TAPANGAN**  
Laboratory Analyst


Reviewed by:

  
**GIDEON M. TANGHAL**  
Laboratory Analyst

Certified Correct and  
Approved for Release by:

  
**JENNIFER J. DEJARME**  
Chief Laboratory Analyst






**TEST REPORT**  
**07-2017-BAL-212Met**

**Sample No.** : Met-450  
**Type of Job** : ON-SITE CALIBRATION  
**Date Calibrated** : July 17, 2017  
**Sample** : Table Balance  
**Manufacturer** : China  
**Serial No.** : 0016  
**Model** : JPT-10  
**Graduation** : 1 g  
**Capacity** : 1000 g  
**Classification of Balance** : Class III (Ordinary Accuracy Class)  
**Company** : **SURIGAO STATE COLLEGE OF TECHNOLOGY**  
**Address** : Narciso Street, Surigao City  
**Page** : 1 of 3


This instrument was calibrated using reference standard traceable to SI Units as maintained by the National Metrology Institute- ITDI, Philippines. The following results were obtained:

**I. Shift Test at 500 g**

Position	Difference (g)	Maximum Permissible Error (MPE)
(1-2)	0.0	$\pm 3.0$ g
(1-3)	0.0	$\pm 3.0$ g
(1-4)	0.0	$\pm 3.0$ g
(1-5)	0.0	$\pm 3.0$ g




OP-026-F14  
Revision 1




**IV. Remarks:**

1. The uncertainty of measurement is estimated at 95% level of confidence with a coverage factor  $k=2$ .
2. The above values are those obtained at the time of test and refer only to the particular instrument calibrated.
3. The end-user shall determine the suitability of this instrument for its intended use.
4. This report shall not be reproduced in any form, except in full, without written approval of the laboratory.

Calibrated by:

  
**ENGR. MANOLITO R. TAPANGAN**  
Laboratory Analyst

Reviewed by:

  
**GIDEON M. TANGHAL**  
Laboratory Analyst

Certified Correct and  
Approved for Release by:

  
**JENNIFER J. DEJARME**  
Chief Laboratory Analyst